SE-85-81 SL-100-822 SL-100-SPG

OTTAWA, June 6, 1969.

## SPECIAL APPROVAL

Granted to: Trans Canada Pipe Lines Limited, 150 Eglinton Avenue East,

Toronto 12, Ontario.

Subject: Kimray Gas Gravitometer,

Serial Number 174150 Instrument Range 0.35 to 1.15 sp. gr.

Special Approval has been granted by the Standards Branch to Trans Canada Pipe Lines Limited for use of the above-named apparatus in connection with billing for gas sales.

This instrument shall be subjected to quarterly re-verification by the Standards Branch.

W. J. J. Fran

W. J. S. Fraser, Chief, Electricity and Gas Division, Standards Eranch.

WJSF/mh

c.c. Mr. J. B. Dumas, D.I of ELG Montreal Mr. Etienne Loranger, Regional Office Montreal