

SPG-26  
SPG-26

SI-5-81  
SI-100-822  
SI-100-SPG

OTTAWA, June 20, 1969.

SPECIAL APPROVAL

Granted to: The Alberta Gas Trunk Line Company Limited,  
505 Second Street S.W.,  
Calgary, Alberta.

Subject: Kinsay Gas Gravitometer,  
Serial Number 174169  
Instrument Range 0.35 to 1.15 sp. gr.

Attention: Mr. C. T. McCall, Measurement Superintendent.

Special approval has been granted by the Standards Branch to the Alberta Gas Trunk Line Company Limited for use of the above-named apparatus in connection with billing for gas sales.

This gravitometer may also be used in the field to spot check gravity on the gas being recorded on field gravitometers.

This instrument shall be subject to quarterly re-verification by the District Office of Electricity and Gas, Operations Branch, Edmonton, Alberta.

W. J. S. Fraser,  
Chief, Electricity and Gas Division,  
Standards Branch.

WJS/wh

c.c. Mr. W. S. Park, D.I. of E&G, Calgary,  
Mr. R. H. McKay, Regional Manager, Winnipeg,  
Mr. H. A. Waterhouse, D.I. of E&G, Edmonton.

