SE-85-81 SL-100-822 SL-100 SPG

CTTAMA, December 2, 1969.

## SPECIAL APPROVAL

Granted to:

Westcoast Transmission Company Ltd.,

1155 West Georgia Street,

Vancouver 5, B.C.

Subject:

Kimray Gas Gravitometer,

Serial Number 174172,

Instrument range 0.25 to 1.16 sp. gr.

Attention:

Er. Don J. Hunter,

Superintendent of Measurement

Special Approval has been granted by the Standards Branch to Westcoast Transmission Company Limited for use of the above-named apparatus in connection with billing for gas sales.

This gravitometer is intended for use in the Fort St. John It may however be used in any area under the jurisdiction of the Edmonton District and shall be subject to quarterly re-verification by that district.

W.J. S. France

W.J.S. Fraser,

Chief,

Electricity and Gas Division.

MULF/oh

c.c. Mr. H. A. haterhouse, D. I. of admonton hr. R. H. McKay, Regional Manager.

....