

SFG-36

SE-85-40-2  
SI-100-71  
SI-100-SFG

OTTAWA, December 22, 1969.

TEMPORARY SPECIAL APPROVAL

Granted to: Northern & Central Gas Corp. Ltd.,  
P.O. Box 1040,  
North Bay, Ontario.

Subject: Canadian Meter Base Volume Index,  
Continuous Integrator - Model 200  
with Printout.

Attention: Mr. A. Brown,  
Measurement Superintendent,  
Northern & Central Gas Corp.

Temporary Special Approval is hereby granted by the Standards Branch  
to Northern & Central Gas Corp. Ltd. to use a Canadian Meter Base Volume  
Continuous Integrator, Model 200, with Print Out in the following installation:

Dupont of Canada Limited - Maitland works  
b.v.i. Continuous Integrator  
-Model 200, printout daily 8.00 A.M.  
Serial No. PF-2-102, Company No. 200370  
Temperature range -20° to 140°F  
Pressure range 0-500 psig  
Base Temp. 60°F  
Atmospheric Pressure 14.48 psia  
Base Pressure 14.73 psia  
Corrected counter reads 1,000 cu. ft.  
Uncorrected counter reads 1,000 cu. ft.

Meter Type  
-Rockwell 8" - 600 ABA Turbo G Meter  
Inspection No. 216006, Working Pressure 225 - 250



Approval for use of this case Volume Index Continuous Integrator  
is temporary pending results of approval tests currently being made.

*W.J.S. Fraser*

W.J.S. Fraser,  
Chief, Electricity and Gas Division

WJF/ah

c.c. Mr. W. Hypponen, Canadian Meter Co. Milton  
Mr. G. O. Piirainen, D.I. of E&G, Sudbury,  
Mr. T. Murphy, D.I. of E&G, Belleville  
Mr. J. W. Kokesh, Regional Manager

