

SPG - 37

SE-85-61  
SL-100-822  
SL-100-SPG

OTTAWA, December 31, 1969.

SPECIAL APPROVAL

Granted to: Trans Canada Pipe Lines Limited,  
150 Kinglinton Avenue East,  
Toronto 12, Ontario.

Subject: Kinray Gas Gravimeter,  
Serial Number 174110,  
Instrument range 0.25 to 1.2 sp. gr.

Special approval has been granted by the Standards Branch to Trans Canada Pipe Lines Limited for use of the above-named apparatus in connection with billing for gas sales.

This gravimeter is intended for use at the Dawn Station in the London Inspection District. It may however be used in any area under the jurisdiction of the London District and shall be subject to quarterly re-verification by that district.

A. J. S. Fraser,  
Chief,  
Electricity and Gas Division.

AJSF/cam

c.c. Mr. J. A. Kosh, Regional Manager, Toronto  
Mr. A. R. Schom, E. I. (120), Toronto  
Mr. J. Fleming, E. I. (120), London.

