

SPG-40

SE-85-40  
SL-100-91  
SL-100-SPG

OTTAWA, January 21, 1970.

SPECIAL APPROVAL

Granted to: Canadian Meter Company Ltd.,  
3037 Derry Road West,  
Milton, Ontario.

Subject: Canadian Meter Base Volume Index  
Continuous Integrator - Model 200  
with Printout.

Attention: Mr. W. Hypponen,  
Manager of Engineering.

Special Approval is hereby granted by the Standards Branch  
to use a Canadian Meter Base Volume Index Continuous Integrator,  
Model 200 with Printout, in the following installations:

1. United Gas Limited, 360 Strathearne North, Hamilton, Ontario.

Model 200, printout daily 8:a.m.  
Serial No. PP-2-031, Company No. 252  
Temperature Range -20 to +140°F  
Pressure Range 0-60 psig  
Base Temperature 60°F  
Atmospheric Pressure 14.40 psia  
Base Pressure 14.65 psia  
Corrected Counter reads 10,000 cu. ft.  
Uncorrected Counter reads 1,000 cu. ft.

Meter Type 250B I.C.  
Serial No. 025094  
Ambient Temperature 60-70°F

*Cancelled  
replaced by SPG-44  
Feb. 12/70*



2. International Harvester Limited, Twine Mill Division, 208 Hillyard Street, Hamilton, Ontario.

Model 200, printout daily 8:a.m.  
Serial No. PP-2-033, Company No. 253  
Temperature Range -20 to +140°F  
Pressure Range 0-60 psig  
Base Temperature 60°F  
Atmospheric Pressure 14.40 psia  
Base Pressure 14.65 psia  
Corrected Counter reads 10,000 cu. ft.  
Uncorrected Counter reads 1,000 cu. ft.

Meter type 500B I.C.  
Serial No. 5055931  
Ambient Temperature 60-70°F

3. Ford Motor Company, Oakville, Ontario.

Model 200, printout daily 8:a.m.  
Serial No. PP-2-034, Company No. 254  
Temperature Range -20 to +140°F  
Pressure Range 0-60 psig  
Base Temperature 60°F  
Atmospheric Pressure 14.40 psia  
Base Pressure 14.65 psia  
Corrected Counter reads 10,000 cu. ft.  
Uncorrected Counter reads 1,000 cu. ft.

Meter type 500B I.C.  
Serial No. 4788433  
To be used outdoors.

The instrument at the Ford Motor Company will be used in a non-billing application but the customer desires government inspection.

*W.J.S. Fraser*

W.J.S. Fraser,  
Chief, Electricity and Gas Division,  
Standards Branch.

WJSF/mh

- c.c. Mr. R. N. Beattie, D.I. of E&G, Hamilton.  
Mr. J. W. Kokesh, Regional Manager, Toronto.  
United Gas Limited, forwarded through Mr. R. N. Beattie.  
International Harvester Limited, forwarded through Mr. R. N. Beattie.  
Ford Motor Company, Oakville, forwarded through Mr. R. N. Beattie.

