

SPG-44

SE-85-40  
SL-100-91  
SL-100-57G

OTTAWA, February 12, 1970.

SPECIAL APPROVAL

Granted to: Canadian Meter Company Ltd.,  
3037 Derry Road West,  
Milton, Ontario.

Subject: Canadian Meter Base Volume Index  
Continuous Integrator - Model 200  
With Printout.

Attention: Mr. W. Hypponen,  
Manager of Engineering.

Special Approval is hereby granted by the Standards Branch to use a Canadian Meter Base Volume Index Continuous Integrator, Model 200 with Printout, in the following installations:

1. United Gas Limited, 360 Strathearne North, Hamilton, Ontario.

Model 200, printout daily 8:a.m.  
Serial No. PP-2-031, Company No. 252  
Temperature Range -20 to +110°F  
Pressure Range 0-60 psig  
Base Temperature 60°F  
Atmospheric Pressure 14.40 psia  
Base Pressure 14.65 psia  
Corrected Counter reads 1,000 cu. ft.  
Uncorrected Counter reads 1,000 cu. ft.

Meter Type 250B I.C.  
Serial No. 025094  
Ambient Temperature 60-70°F



2. International Harvester Limited, Twino Mill Division 208 Hillyard Street, Hamilton, Ontario.

Model 200, printout daily 8:a.m.  
Serial No. PP-2-033, Company No. 253  
Temperature Range -20 to +140°F  
Pressure Range 0-60 psig  
Base Temperature 60°F  
Atmospheric Pressure 14.40 psia  
Base Pressure 14.65 psia  
Corrected Counter reads 1,000 cu. ft.  
Uncorrected Counter reads 1,000 cu. ft.

Meter type 500 B I.C.  
Serial No. 5055931  
Ambient Temperature 60-70°F

3. Ford Motor Company, Oakville, Ontario.

Model 200, printout daily 8:a.m.  
Serial No. PP-2-034, Company No. 254  
Temperature Range -20 to +140°F  
Pressure Range 0-60 psig  
Base Temperature 60°F  
Atmospheric Pressure 14.40 psia  
Base Pressure 14.65 psia  
Corrected Counter reads 1,000 cu. ft.  
Uncorrected Counter reads 1,000 cu. ft.

Meter type 500 B I.C.  
Serial No. 4732133  
To be used outdoors.

The instrument at the Ford Motor Company will be used in a non-billing application but the customer desires government inspection.

M. J. S. Fraser,  
Chief,  
Electricity and Gas Division,  
Standards Branch.

c.c. Mr. R. N. Beattie, D.I. of E&G, Hamilton  
Mr. J. W. Kokesh, Regional Manager, Toronto.

NOTE: Special Approval SPG-40 cancelled and replaced by SPG-44.

