

OTTAWA, February 20/70

SPG-45  
11-15

SI-100-02  
SI-100-01  
SI-100-01G

SPECIAL APPROVAL

Granted to: Saskatchewan Power Corporation,  
Regina, Sask.

Subject: Canadian Meter Base Volume Index Continuous  
Integrator - Model 100 without Printout

Attention: Mr. D. W. Campbell,  
Technical Services Superintendent

Special approval is hereby granted by the Standards Branch to Saskatchewan Power Corporation to use the following instrument in the measurement of gas:

- Canadian Meter Base Volume Index, Continuous Integrator
- Model 100 without printout
  - Serial No. 11-1-002
  - Inspection No. 200
  - Pressure 0-100 psig
  - Atmospheric Pressure 13.65 psig
  - Base Pressure 13.9 psia
  - Temperature -40°F to +140°F
  - Base Temperature 60°F
  - Corrected counter x 1000 in cubic feet
  - Uncorrected counter x 100 in cubic feet
  - Supercompressibility factor not included
  - for use on a 100 ft. drive meter.

J.F./A

W.J.S. Fraser,  
Chief, Electricity and Gas Division

c.c Mr. J. Luff, D.I. of E&G Regina  
Mr. R. H. Ackay, Regional Manager, Winnipeg  
Mr. R. M. Beattie, D.I. of E&G Hamilton

