



Department of consumer and corporate affairs / Ministère de la consommation et des corporations



SPG-47
SE-85-G
SL-100-G
SL-100-SPG

OTTAWA, May 22, 1970.

SPECIAL APPROVAL

Granted to: Saskatchewan Power Corporation,
Regina, Saskatchewan.

Subject: Varian Aerograph Gas Chromatograph
Part No. 1720-1, Serial #02760823

Attention: Mr. D.F. Campbell,
Technical Services Superintendent.

Special Approval has been granted by the Standards Branch to the Saskatchewan Power Corporation for the use of the above named apparatus for the determination of the B.t.u. and Specific Gravity values of natural gas. The determination depends upon the use of continuous gas sampling devices and the chromatographic analysis of the sample gases thus collected.

A check of the apparatus shall be made every three months by the Regina District Office. Such check is based upon the B.t.u. and Specific Gravity values determined by chromatographic analysis of the District Office standard gas compared to its certified values.

W.J.S. Fraser,
Chief, Electricity and Gas Division.

cc: Mr. J. Duff, D.I. E&G, Regina
Mr. R.H. McKay, Regional Manager, Winnipeg.

