

SPG-50
~~SP-50~~

SI-100-607
SI-89-58
SI-100-SPG

OTTAWA, May 5, 1970.

SPECIAL APPROVAL

Granted to: The Alberta Gas Trunk Line Co. Ltd.,
505 Second Street S.W.,
Calgary, Alberta.

Subject: Modified Cutler-Hammer Type AB
Calorimeters

Attention: Mr. C. T. McCall,
Measurement Superintendent

Special Approval has been granted by the Standards Branch to the Alberta Gas Trunk Line to modify its Cutler-Hammer Type AB Calorimeters by replacing the Leeds and Northrup Micromax recorders with Leeds and Northrup Speedomax Type W Strip Chart Recorders.

The modification procedure applied to all units shall be in essential accord with that carried out on Cutler-Hammer Calorimeter Type AB, Serial No. 0963263A.

The Type W Leeds and Northrup Recorder used in this modification is similar to those approved under Approval Notice E-53, May 4, 1966, except that it incorporates a Wheatstone Bridge instead of a potentiometer as the measuring element.

M. J. S. Fraser,
Chief, Electricity and Gas Division.

c.c. Mr. A. S. Park, D.I. of E&G, Calgary
Mr. A. H. McKay, Regional Manager, Winnipeg, Manitoba.

