

CIG-51

41-5-50
70-100-100

OTTAWA, June 3, 1970.

SPECIAL APPROVAL

Granted to: Union Gas Co. (Canada) Ltd.,
Chatham, Ontario.

Subject: Rockwell Type T-140-10" 300 ANSI Turbo-meters.

Attention: Mr. J.A. Reid,
Union Gas Co.

Special approval has been granted by the Standards Branch to Union Gas Company for an installation, incorporating two Type T-140-10" 300 ANSI Turbo-meters, at Polymer Corporation, Vidal Street, Sarnia. The serial numbers of these meters are 1030012 and 1030013.

The meters are each fitted with Rockwell Type T transmitters, inspection numbers 8V01126 and 8V01127 respectively, and serial numbers 83339 and 83340 respectively. The ranges and other basic data for the instruments are as follows:

- Base Pressure	14.65 psig
- Base Temperature	60°F
- Pressure range	0-500 psig
- Temperature range	-20° to 150°F
- Drive	1000 c.f./rev.

W.J.B. Fraser,
Chief,
Electricity and Gas Division,
Standards Branch.

cc: Rockwell Manufacturing Co. (Canada) Ltd., Guelph, Ont.
Attention: Mr. R.A. Laidman, General Manager, Gas Div.

Mr. W.A. McEwen, Reg. Supervisor E&G Toronto
Mr. J.W. Menezes, Reg. Manager, Toronto
Mr. J. Fleming, Dist. Insp. E&G, London, Ont.

Attached Copy of Rockwell letter &

EG-1 dwgⁿ - SE-85-41

