

SFC-52
SL-65-62
SL-100-240
SL-100-SFC

OTTAWA, May 29, 1970.

SPECIAL APPROVAL

Granted to: Trans Canada Pipelines Ltd.,
150 Wellington Avenue East,
Toronto 12, Ontario.

Subject: Modified Cutler-Hammer Type AB Calorimeter

Attention: Mr. P. Artemchuk
Gas Measurement,
Trans Canada Pipelines Limited.

Special Approval has been granted by the Standards Branch to Trans Canada Pipelines Ltd., to modify a Cutler-Hammer Type AB Calorimeter by replacing the Leeds and Northrup Micromax recorder with a Leeds and Northrup Speedomax Type W Strip Chart Recorder.

The calorimeter is identified as #11368 serial #791649A. The Speedomax Type W recorder is identified as Cat. #500-911-000-0999-002-6-006-005-9, Serial number C-69-1753-001-1-1.

The Type W Leeds and Northrup Recorder used in this modification is similar to those approved under Approval Notice E-33, May 4, 1966, except that it incorporates a Wheatstone Bridge instead of a Potentiometer as the measuring element.

The calorimeter is located at TCPL Station 130.

W.J.S. Fraser,
Chief,
Electricity and Gas Division,
Standards Branch.

cc: Mr. J.W. Kokesh, Reg. Manager, Toronto
Mr. W.R. McIown, Reg. Supervisor, Toronto
E&G

