



SPG-57

SE-85-81
SL-100-822
SL-100-SPG

OTTAWA, June 23, 1970.

SPECIAL APPROVAL

Granted to: Trans Canada Pipe Lines Limited,
150 Eglinton Avenue East,
Toronto 12, Ontario.

Subject: Kimray Gas Gravitometers
Serial Nos. 174120 and 174140
Instrument range: 0.50 to 1.0 (same for both).

Attention: Mr. G. B. Miller

Special Approval has been granted by the Standards Branch for use of the above named apparatus in connection with billing for gas sales.

The gravitometers are intended for use in areas as follows:

Serial No. 174120	Maple, Ontario
Serial No. 174140	Grimsby, Ontario.

These gravitometers may also be used in other areas under the jurisdiction of the respective Districts and they shall be subject to quarterly re-verification by our field inspector.

(for) W.J.S. Fraser,
Chief,
Electricity and Gas Division.

c.c. Mr. W. R. McEown, Regional Supervisor, E&G, Toronto
Mr. R. N. Beattie, D.I. of E&G, Hamilton.

