



SPG-60
SE-85-56
SL-100-947
SL-100-SPG

OTTAWA, July 15, 1970.

SPECIAL APPROVAL

Granted to: Union Gas Company (Canada) Ltd.
Chatham, Ontario.

Subject: Parkinson & Cowan Delta N1 Rotary Meter
Max. Working Pressure 720 psig, 7000 cfh capacity

Attention: Mr. J. A. Reid,
Union Gas Company (Canada) Ltd.

Special Approval has been granted by the Standards Branch to Union Gas Company for the installation of one high pressure Delta N1 Rotary Meter, serial number 126020 in Somba Township, Lambton County, Lot 26, Concession 7.

This meter will be equipped with

- (a) an approved Canadian Meter Volume and Pressure gauge, 0-500 psi, with flexible drive cable
- (b) an approved Foxboro temperature recorder, 0-150°F.

The average working pressure at the installation is expected to be in the neighborhood of 400 psig.

W.J.S. Fraser,
Chief, Electricity and Gas Division

c.c. Parkinson Cowan (Canada) Ltd.
Attention: Mr. H. Bennett, Sales Representative
Mr. W. R. McEown, Regional Supervisor E&G Toronto
Mr. J. Fleming, D.I. of E&G London

