

SPG-61
SE-85-81
SL-100-822
SL-100-SPG

OTTAWA, July 21, 1970.

SPECIAL APPROVAL

Granted to: The Alberta Gas Trunk Line Company Ltd.,
505 Second Street S.W.,
Calgary, Alberta.

Subject: Kinray Gas Gravitometers,
Serial Nos. 174187 and 174189,
Range of each 0.35 to 1.15

Attention: Mr. C. T. McCall,
Measurement Superintendent,
Alberta Gas Trunk Line Company.

Special Approval has been granted by Standards Branch to the Alberta Gas Trunk Lines Company Limited for the use of the above-named instruments in connection with billing for gas sales.

These gravitometers will be used out of the company Service Centre for checking field gravitometers. However, they may be used in any area under the jurisdiction of the Calgary District Office and shall be subject to quarterly re-verification by that District.

W.J.S. Fraser
W.J.S. Fraser,
Chief,
Electricity and Gas Division.

c.c. Mr. A. S. Park, D.I. of E&G, Calgary
Mr. R. H. McKay, Regional Manager, Winnipeg.

