

SPG-69

SE-85-40
SL-100-91
SL-100-SPG

OTTAWA, August 24, 1970

SPECIAL APPROVAL

Granted to: Kitchener P.U.C.,
191 King Street, West,
Kitchener, Ont.

Subject: Canadian Meter Base Volume Index Continuous
Integrator - Model 200

Attention: Mr. J. Kranenburg

Special Approval is hereby granted by the Standards Branch
to Kitchener P.U.C. to use the following instrument in the measure-
ment of gas:

Canadian Meter Base Volume Index, Continuous Integrator
Instrument Location - B. F. Gooderich Plant
-Model 200

- Serial No. PP 2-144
- Company No. 53
- Temperature Range -20° to 140°F
- Pressure Range - 0-45 PSIG
- Base Temperature 60°F
- Atmospheric Pressure 14.4 PSIA
- Base Pressure 14.65 PSIA
- Corrected Counter reads 1,000 cu. ft.
- Uncorrected Counter reads 100 cu. ft.
- No supercompressibility factor
- To be used on A.M. 6" Turbine

W.J.S. Fraser

W.J.S. Fraser,
Chief, Electricity and Gas Division.

c.c. Mr. R. N. Beattie, Hamilton
Mr. J. W. Ostler, Milton
Mr. J. A. Orwin, Milton

