



SPG-70

SE-85-40
SL-100-91
SL-100-SPG

OTTAWA, August 24, 1970.

SPECIAL APPROVAL

Granted to: Alberta Natural Gas Company,
240-4th Avenue S.W.,
Calgary 1, Alberta

Subject: Canadian Meter Base Volume Index Continuous
Integrator - Model 200

Attention: Mr. M. Winton

Special Approval is hereby granted by the Standards Branch to Alberta Natural Gas Company to use the following instrument in the measurement of gas:

- Canadian Meter Base Volume Index, Continuous Integrator
Instrument Location - Crows Nest Station #1
- Model 200
 - Serial No. PP - 2- 015
 - Company No. 5
 - Temperature Range -20° to +140°F
 - Pressure Range 0-1000 PSIG
 - Base Temperature 60°F
 - Atmospheric Pressure 13.37 PSIA
 - Base Pressure 14.73 PSIA
 - Corrected counter reads 10,000 cu. ft.
 - Uncorrected counter reads 1,000 cu. ft.
 - No supercompressibility factor
 - To be used on Rockwell T-18
 - Inside location at +50°F

W.J.S. Fraser
W.J.S. Fraser,
Chief, Electricity and Gas Division.

c.c. Mr. R. N. Beattie, Hamilton
Mr. J. W. Ostler, Milton
Mr. J. A. Orwin, Milton
Mr. A. S. Park, Calgary
Mr. J. M. Taylor, Edmonton

Standards Branch
Direction des normes

