

SPG-71

SE-85-40
SL-100-91
SL-100-SPG

OTTAWA, August 28, 1970.

SPECIAL APPROVAL

Granted to: United Gas Company,
Hamilton, Ontario.


Subject: One Canadian Meter Base Volume Index
Continuous Integrator - Model 200
With Printout.

Attention: Mr. H. W. Hyslop,
General Superintendent.

Special Approval is hereby granted by the Standards Branch to United Gas Company to use the following instrument in the measurement of gas:

Canadian Meter Base Volume Index, Continuous Integrator Model
200
Instrument Location - Woodward Ave. Sewage Disposal Plant in
Hamilton

- Model 200
- Serial No. PP - 2 - 037
- Company No. 256
- Temperature Range -20° to +140°F
- Pressure Range 0-60 PSIG
- Base Temperature 60°F
- Atmospheric Pressure 14.4 PSIA
- Base Pressure 14.65 PSIA
- Corrected Counter reads - 1,000
- Uncorrected Counter reads - 1,000
- No supercompressibility factor
- To be used on 8" Rockwell Series T-60 Turbo-Meter



W.J.S. Fraser,
Chief, Electricity & Gas Division.

c.c. Mr. R. N. Beattie, Hamilton
Mr. J. A. Orwin - Milton
Mr. J. W. Ostler - Milton

Standards Branch
Direction des normes

