



SPG-73

SE-85-40
SL-100-91
SL-100-SPG

OTTAWA, August 31, 1970.

SPECIAL APPROVAL

Granted to: United Gas Company,
Hamilton, Ontario.

Subject: One Canadian Meter Base Volume Index
Continuous Integrator - Model 200
With Printout

Attention: Mr. H. W. Hyslop,
General Superintendent.

Special Approval is hereby granted by the Standards Branch to United Gas Company to use the following instrument in the measurement of gas:

- Canadian Meter Base Volume Index, Continuous Integrator Model 200
- Instrument Location - Ford Motor Company, Oakville, Ontario
- Model 200, print out daily 8: a.m.
- Serial No. PP - 2 - 034
- Company No. 254
- Temperature Range -20°F to +140°F
- Pressure Range 0-60 PSIG
- Base Temperature 60°F
- Atmospheric Pressure 14.40 PSIA
- Base Pressure 14.65 PSIA
- Corrected counter reads 1,000 cu. ft.
- Uncorrected counter reads 1,000 cu. ft.
- No Supercompressibility Factor
- To be used on a Canadian 8" GT - 60M Turbine Meter
- To be used outdoors.

This approval cancels out that which was granted under SPG-44.

W. J. S. Fraser

W.J.S. Fraser,
Chief, Electricity & Gas Division.

c.c. Mr. R. N. Beattie, Hamilton
Canadian Meter Company, Milton

Standards Branch
Direction des normes

