

SPR-73

SL-100-100

SL-100-100

SL-100-SPC

OTTAWA, September 21, 1970.

SPECIAL APPROVAL

Granted to: The Alberta Gas Trunk Line Company Ltd.,
P.O. Box 1599,
Calgary 2, Alberta.

Attention: Mr. C. T. McGill,
Measurement Superintendent.

Subject: Airway Gas Gravimeter
Serial No. 174185
Range 0.35 to 1.15

Special Approval has been granted by Standards Branch to the Alberta Gas Trunk Lines Company Limited for the use of the above-named instrument in connection with billing for gas sales.

This gravimeter will be used out of the company Service Centre for checking field gravimeters. However, it may be used in any area under the jurisdiction of the Calgary District Office and shall be subject to quarterly re-verification by that District.

A.J.S. Fraser,
Chief,
Electricity and Gas Division.

c.c. Mr. A. S. Park - Calgary
Mr. R. H. McKay, Regional Manager, Winnipeg.

