

SPG-102

SL-100-667

SL-69-C

SL-100-SPG

OTTAWA, February 18, 1971.

SPECIAL APPROVAL

Granted to: Westcoast Transmission Co. Ltd.,
1133 West Georgia Street,
Vancouver, B.C.

Attention: Mr. D. J. Hunter,
Superintendent of Measurement

Subject: Modified Cutler-Hammer Type B Calorimeter

Special Approval has been granted by the Standards Branch to Westcoast Transmission Company to modify a Cutler-Hammer Type AB Calorimeter by replacing the Leeds and Northrup Micromax Recorder with a Leeds and Northrup Type W Strip Chart Recorder.

The Recording Calorimeter is installed at the Huntingdon meter station and has the Serial No. 915896A. The new Type W recorder has Serial No. C70-2496-C01-1-1, range 900-1200 B.T.U., entrance thermometer Number 6130A, exit thermometer Number 6130B, and slide-wire Number 6130.

The Type W Leeds and Northrup Recorder used in this modification is similar to those approved under Approval Notice E-33, May 4, 1966, except that it incorporates a Wheatstone Bridge instead of a Potentiometer as the measuring element.

W. J. S. Fraser

W.J.S. Fraser,
Chief,
Electricity and Gas Division.

c.c. Mr. R.G.D. Ward, D.I. of E&G, Vancouver

