

SPG-113

SE-85-44
SL-100-855
SL-100-SFG

OTTAWA, June 1, 1971.

SPECIAL APPROVAL

Granted to: Greater Winnipeg Gas Company,
265 Notre Dame Ave.,
Winnipeg, Manitoba.

Attention: Mr. W. Ferguson,
Superintendent of Operations.

Subject: Roots Model 7M1440 High Pressure Rotary Meter

Special Approval has been granted by the Standards Branch to Greater Winnipeg Gas Limited for the installation of one Roots Model 7M1440 high pressure meter rated at 1440 psig.

Details of the meter are:

Company No.	134490
Serial No.	7108991
Meter Pressure - Max.	800 psig
- Min.	350 psig
- Normal	500 psig
Instrument Drive	10 cu. ft./rev.

Meter equipped with Canadian Meter BVI Continuous Integrator Model 100 without printout

Pressure Range	0-1000 psi
Temperature Range	-20° to +140°F
Base Temperature	60°F
Base Pressure	14.55 psi
Atmospheric Pressure	14.30 psi

Supercompressibility factor included based on 500 psig, 50°F
Multiplier for uncorrected counter 100
Multiplier for corrected counter 1000

*See handwritten notes below
Bill says copy 26 271
file 85-42 21
24-150-101*

J. S. Fraser
(for) W.J.S. Fraser,
Chief,
Electricity and Gas Division

c.c. Mr. D.E. Kohart, Dresser Industries
Mr. N. H. Young, D.I. of E&G Winnipeg
Mr. W. R. McEown, D.I. of E&G Toronto
Mr. R. N. Beattie, D.I. of E&G Hamilton

