



Department of consumer and corporate affairs, Ministère de la consommation et des corporations



SPG-139

1147-57/R2-M29

SL-100-978 D

SL-100-SPG

OTTAWA, November 26, 1971.

SPECIAL APPROVAL

Granted to: Columbia Natural Gas Ltd.,  
Cranbrook, B.C.

Attention: Mr. K. McLean,  
General Manager.

Subject: Two Rockwell Turbo-meters Series G  
Model T-60 Mark II

Special Approval has been granted by the Standards Branch to Columbia Natural Gas for the use of the following equipment in the measurement of gas:

Rockwell Series G 8" Turbomotors  
Models T-60 Mark II 60,000 cfh capacity  
Company No. CNG 7335 Serial G810000  
Company No. CNG 7336 Serial G810001

These meters are aluminum case and are rated 125 psi rated working pressure.

*W. J. S. Fraser*

W. J. S. Fraser,  
Chief,  
Electricity and Gas Division.

c.c. Mr. A. L. Heaven, Penticton, B.C.  
Mr. R.G.D. Ward, Vancouver, B.C.  
Mr. W. R. McEown, Toronto, Ont.  
Rockwell Manufacturing Co., Guelph, Ont.

