



Department of consumer and corporate affairs / Ministère de la consommation et des corporations

SPG-140

1147-57/K3
SL-100-822
SL-100-SPG

OTTAWA, December 14, 1971.

SPECIAL APPROVAL

Granted to: The Alberta Gas Trunk Line Company Ltd.,
P.O. Box 2535,
Calgary 2, Alberta.

Attention: Mr. T. N. Befus, P. Eng.,
Measurement Superintendent.

Subject: Two Kimray Gas Gravitometers

Special Approval has been granted by Standards Branch to the Alberta Gas Trunk Lines Company Limited for the use of the above-named instrument in connection with billing for gas sales.

Specifications on the two units are as follows:

Calgary: Serial No. 174193	Edmonton: Serial No. 266686
A.G.T.L. No. 3-01-07	A.G.T.L. No. 3-01-01
Range 0.35 to 1.15	Range 0.35 to 1.15
Power Supply 120 Volts A.C.	Power Supply 120 Volts A.C.

These Gravitometers will be used out of the company Service Centres for checking field gravitometers. However they may be used in any area under the jurisdiction of the Calgary and/or Edmonton District Office and shall be subject to quarterly re-verification by that District.

W. J. S. Fraser

W. J. S. Fraser,
Chief,
Electricity and Gas Division.

c.c. Mr. A. S. Park, Calgary
Mr. J. Taylor, Edmonton

