



SPG-150

1147-57/K3
SL-100-822
SL-100-SPG

OTTAWA, March 13, 1972.

SPECIAL APPROVAL

Granted to: The Alberta Gas Trunk Line Company Ltd.,
P.O. Box 2535,
Calgary 2, Alberta.

Attention: Mr. T. N. Befus, P. Eng.,
Measurement Superintendent.

Subject: Two Kimray Gas Gravitometers

Special Approval has been granted by Standards Branch to the Alberta Gas Trunk Lines Company Limited for the use of the above-named instruments in connection with billing for gas sales.

Specifications on the two units are as follows:

- | | |
|-----------------------------|-----------------------------|
| 1. Serial No. 277611 | 2. Serial No. 277615 |
| A.G.T.L. No.3-01-14 | A.G.T.L. No.3-01-15 |
| Range - 0.35 to 1.15 | Range - 0.35 to 1.15 |
| Power Supply 120 Volts A.C. | Power Supply 120 Volts A.C. |

These Gravitometers will be used out of the company Service Centres for checking field gravitometers. However they may be used in any area under the jurisdiction of the Calgary District Office and shall be subject to quarterly re-verification by the Calgary District.

W.J.S. Fraser,
Chief,
Electricity & Gas Division.

c.c. Mr. A. S. Park, Calgary

