



SDG-151

1147-57/K3
SL-100-322
SL-100-SPG

OTTAWA, March 24, 1972.

SPECIAL APPROVAL

Granted to: The Alberta Gas Trunk Line Company Ltd.,
P.O. Box 2535,
Calgary 2, Alberta.

Attention: Mr. T. N. Befus, P. Eng.,
Measurement Superintendent.

Subject: One Kimray Gas Gravitometer

Special Approval has been granted by Standards Branch to the Alberta Gas Trunk Line Company for the use of the above-named instrument in connection with billing for gas sales.

Specification on the unit is as follows:

Serial No.	277618
A.G.T.L. No.	3-01-16
Range	0.35 to 1.15
Power Supply	120 Volts A.C.

The gravitometer will be used to spot check the specific gravity of gas recorded on field gravitometers. It may be used in any area under the jurisdiction of the Calgary District Office and shall be subject to quarterly re-verification by the Calgary District.

W.J.S. Fraser
W.J.S. Fraser,
Chief,
Electricity and Gas Division.

cc Mr. A. S. Park, Calgary.

