




Department of consumer and corporate affairs / Ministère de la consommation et des corporations



SPG-153  
1147-57/R2-M29  
SL-100-981  
SL-100-SFG

OTTAWA, April 17, 1972

SPECIAL APPROVAL

Granted to: Gaz Metropolitan Inc.,  
1717 du Havre,  
Montreal 134, P.Q.

Attention: Mr. A. Lodomez,  
Meter Shop

Subject: Fisher and Porter Model 1543A13A-NS33  
Gas Survey Recorder with Rockwell T-60  
Turbo meter fitted with Rockwell Model  
3550 Rotaswitch contactor

Special Approval has been granted by the Standards Branch to Gaz Metropolitan to apply the subject instrumentation in one installation the details of which are as follows:

1. Labatt Brewries,  
50 Labatt Street,  
Ville LaSalle, Que.
  - (a) Fisher & Porter gas survey recorder  
Model 1543A13A-NS33  
Serial No. 7203B2186K1  
1 pulse records 1000 cu. ft. corrected
  - (b) Rockwell temcorector Type I - 0-100 psig range  
Serial No. 8-7257  
Inspector No. 96-149381  
1 revolution drive = 1000 cu. ft.
  - (c) Impulse Contactor - Disc Rotaswitch  
Model 3550  
1 pulse - 1000 cu. ft. (corrected)

In addition to the test normally required for Temcorectors and Survey Recorders, test and inspection shall be made with the



appropriate components properly connected together and performing as an entity.

*W.J.S. Fraser*

W.J.S. Fraser,  
Chief,  
Electricity and Gas Division,  
Standards Branch.

c.c. Mr. A. Laferriere, E&G, Montreal

