




Department of consumer and corporate affairs / Ministère de la consommation et des corporations



SPG-155

SL-100-822

1147-4

OTTAWA, April 19, 1972

SPECIAL APPROVAL

Granted to: Johnston Testers,
321-50th Avenue S.E.,
Calgary 24, Alberta

Attention: Mr. G. A. Welsh,
Manager-Production Services

Subject: One Kimray Gas Gravitometer

Special Approval has been granted by Standards Branch to Johnston Testers for the use of the above-named instrument in connection with billing for gas sales.

Specification on the unit is as follows:

Serial No.	277609
Range	0.35 - 1.15 sp. gr.
Power Supply	120 volts A.C.

The gravitometer will be used to spot check the specific gravity of gas recorded on field gravimeters. It may be used in any area under the jurisdiction of the Calgary District Office and shall be subject to quarterly re-verification by the Calgary District.

W. J. S. Fraser

W.J.S. Fraser,
Chief,
Electricity and Gas Division.

c.c. Mr. A. S. Park, Calgary

