



SPG-164

SI-100-85
1117-50/85-164
SI-100-85

OTTAWA, May 25, 1972

SPECIAL APPROVAL

Granted to: United Gas Limited,
 360 Strathcarne Ave. E.,
 Hamilton, Ontario.

Attention: Mr. Jerry Peters

Subject: Parkinson Cowan - 4", 14,000 cfm "Fluxi"
 with instrument drive.

Special Approval has been granted by the Standards Branch to United Gas Limited for the installation of one 4", 14,000 cfm, Compagnie Des Compteur "Fluxi" Turbine Gas Meter at Long Manufacturing Co. Ltd., 656 Kerr Street, Oakville, Ont. The meter is fitted with an instrument drive.

Meter: 14,000 "Fluxi" - 4"
 Serial No.: 315249
 Inspection No.: 35942
 Average Working Pressure: 40 psig

Instrument: Canadian Meter VPT .
 pressure range 0-100 psi
 temperature range - 32°F. to 120°F.

W.J.S. Fraser
Chief,
Electricity and Gas Division

