

SPG-183
JL-180-981
1147-57/82-129
1147-4

OTTAWA, July 31, 1972

TEMPORARY SPECIAL APPROVAL

Granted to: Gaz Metropolitain Inc.,
1717 du Havre,
Montreal 134, P.Q.

Attention: Mr. A. Lodomez,
Meter Shop.

Subject: Fisher and Porter Model 1543-FA13A Gas
Survey Recorder with Rockwell Type I
Temporector fitted with a Rockwell Model
3550 Rota Switch Contactor.

Special Approval has been granted by the Standards Branch to Gaz Metropolitain to apply the subject instrumentation in one installation the details of which are as follows:

1. Labatt Brewries Ltd.,
50 Labatt Street,
Ville LaSalle, Que.
 - (a) Fisher & Porter Recorder
Model 1543 - FA13A
Serial No. 7202B2309K1
Inspection No. 184-197275
1 pulse recorder = 10,000 cu. ft.
 - (b) Rockwell Temporector Type I
Range 0-20 psig
Serial No. 1G-5014
Inspection No. 96-197290

(c) Impulse Contactor
Type Disc Rota Switch
Model No. 3550
1 Pulse - 1000 cu. ft. (corrected)

(d) Meter
Rotary Type 16M125
Inspection No. 112-90627
Serial No. 6405272

In addition to the test normally required for Temcorectors and Survey Recorders, test and inspection shall be made with the appropriate components properly connected together and performing as an entity.

W.J.S. Fraser,
Chief,
Electricity and Gas Division.

/ c.c. Mr. J. B. Dumas

