

SPG-184
SL-100-981
1147-57/R2-1129
1147-4

OTTAWA, July 31, 1972.

TEMPORARY SPECIAL APPROVAL

Granted to: Gaz Métropolitain Inc.,
1717 du Havre,
Montreal 134, P.Q.

Attention: Mr. A. Lodomez,
Meter Shop.

Subject: Fisher and Porter Model 1543-FA13A Gas
Survey Recorder with Rockwell Type I
Temporector fitted with a Rockwell Model
3550 Rota Switch Contactor.

Special Approval has been granted by the Standards Branch to Gaz Métropolitain to apply the subject instrumentation in one installation the details of which are as follows:

1. Maple Leaf Ceramics,
1740 Bois Franc,
Ville St. Laurent, P.Q.
 - (a) Fisher & Porter Gas Survey Recorder
Model 1543FA13-NS33
Serial No. 7202E-2308K1
Inspection No. 184-197274
Recording 1 pulse 1000 cu. ft.
(corrected)
 - (b) Rockwell Temporector Type 1
Inspection No. 96-197291
Serial No. 1G-5015
Range 0-20 psig
1 Rev. 1000 cu. ft. (corrected)

- (c) Impulse Contactor
Type Disc. Rota Switch
Model No. 3550
1 revolution geared magnet = 10,000 corrected
cu. ft.

- (d) Meter
Make-Roots 8 x 24 Rotary
Inspection No. 75-45130
Serial No. 25320
1 Rev. output = 1,000 cu. ft.
Uncorrected.

In addition to the test normally required for Temcorectors and Survey Recorders, test and inspection shall be made with the appropriate components properly connected together and performing as an entity.

W.J.S. Fraser
W.J.S. Fraser,
Chief,
Electricity and Gas Division.

c.c. Mr. J.B. Dumas

