

SPG-188
~~SPG-188~~
1147-57/A2-98/
SL-100- 98/
1147-4

Ottawa, August 15, 1972

SPECIAL APPROVAL

Granted to: Gaz Metropolitan Inc.,
1717 du Havre,
Montreal 134, P.Q.

Attention: Mr. A. Lodomez,
Meter Shop.

Subject: Fischer and Porter Model 1543-PA-13-NS33
Gas Survey Recorder with Rockwell Type I
Temporecator fitted with a Rockwell Model
3550 Rota Switch Contactor.

Special Approval has been granted by the Standards Branch
to Gaz Metropolitan to apply the subject instrumentation in
one installation the details of which are as follows:

1. Continental Can Company Ltd.,
4455 Cote de Liesse,
St. Laurent, P.Q.
 - (a) Fischer and Porter gas survey recorder
Model 1543-PA-13-NS33
Serial Number 7205B2108K4
Inspection Number 184-197301
1 pulse records 1000 cu. ft. (corrected).
 - (b) Rockwell temporecator Type I - 0-20 psig range
Inspection Number 96-197-290
Serial Number IC - 5014
1 revolution drive = 100 cu. ft.
 - (c) Impulse contactor - Disc Rota Switch
Model 3550
1 pulse - 1,000 cu. ft (corrected)

In addition to the test normally required for Temporecators
and Survey Recorders, test and inspection shall be made with the

appropriate coal assets properly connected together and performing as an entity.

H. H. PRASOR,
Chief,
Electricity and Gas Division,
Standards Branch.

c.c. Mr. J. E. Dumas, Montreal

