



SPG-194

Department of consumer and corporate affairs, Ministère de la consommation et des corporations

SPG-194

1147-1

SI-100-901

1147-57/82-90

Ottawa, October 4, 1972.

SPECIAL APPROVAL

Granted to: Gaz Metropolitan, Inc.,  
1717 de Bayre,  
Montreal 134, P.Q.

Attention: Mr. A. Ledomez,  
Meter Shop

Subject: Fisher and Porter Model 1543 Gas Survey  
Recorder with Rockwell Type I Temcorator  
fitted with a Rockwell Model 3550 Rota Switch  
Contactor.

Special Approval has been granted by the Standards Branch to Gaz Metropolitan to apply the subject instrumentation in one installation the details of which are as follows:

1. Canadian Steel Wheel Limited,  
1900 Dickson Street,  
Montreal, P.Q.
  - (a) Fisher and Porter Gas Survey Recorder  
Model 1543  
Serial No. 7207B2009X1  
Inspection No. 184-197355  
1 pulse records 1000 cu. ft. (corrected)
  - (b) Rockwell Temcorator Type I - 0-100 psi range  
Inspection No. 96-197345  
Serial No. 165019  
1 revolution drive = 100 cu. ft. (corrected)  
Pulse Disc Rota Switch No. 3550
  - (c) Meter  
Roots Rotary 38M125  
Inspection No. 114-197281  
Serial No. 7113185  
Pressure 70 psi



In addition to his usual monthly reports for Gasconnectors and Surveying, connecting work and inspection shall be made with the appropriate Gas Connectors, properly connected together and performing as an entity.

H.T.S. STANCO,  
Chief, Electricity and Gas Division,  
Standards Branch.

C.C. District Inspector (Hortland)  
The J. B. James

