



SPG-209

C 1147-7  
GL 1147-7

OTTAWA, January 9, 1973.

SPECIAL APPROVAL

Granted to: Gaz Métropolitain Inc.,  
1717 du Havre,  
Montreal 134, P.Q.

Attention: Mr. A. Lodomez,  
Meter Shop.

Subject: Fischer & Porter Gas Survey Recorder  
with Rockwell Type I Temcorector fitted  
with a Rockwell Disc Rota Switch

Special Approval has been granted by the Standards Branch to Gaz Métropolitain to apply the subject instrumentation in one installation, the details of which are as follows:

Canadian Titanium & Pigments,  
Varenes,  
Quebec.

- |     |  |   |
|-----|--|---|
| (a) | Fischer & Porter<br>Inspection No.<br>Serial No.<br>Recording  | Gas Survey<br>184-149390<br>6911-B2010-P1<br>1 pulse = 1000 <sup>3</sup> ft.  |
| (b) | Rockwell Temcorector<br>Type<br>Inspection No.<br>Serial No.<br>Range<br>Drive<br>Corrected reading<br>Signal generator<br>1 pulse | 1<br>96-149377<br>7255<br>0-100 psig<br>100 <sup>3</sup> ft. rev.<br>1000 <sup>3</sup> ft. increment<br>Disc Rota Switch<br>1000 <sup>3</sup> ft. |
| (c) | Roots Rotary<br>Inspection No.<br>Serial No.<br>Instrument Drive per<br>rev.<br>Inlet pressure                                     | 10 x 30<br>76-49987<br>25579<br>100 <sup>3</sup> ft.<br>30 psig   |

In addition to the test normally required for  
Temcorectors and Survey Recorders, test and inspection  
shall be made with the appropriate components properly  
connected together and performing as an entity.

W.J.S. Fraser,  
Chief,  
Electricity and Gas Division.

c.c. District Inspector Montreal  
Mr. J. B. Dumas