



OTTAWA, Ontario
April 2, 1974

Your file *voire référence*

Our file *Notre référence*

GL 1147-57/C6-123
G 1147-57/C6-123
GL 1147-57/M7-107
G 1147-57/M7-107

SPECIAL APPROVAL

Granted to: The Montana Power Company,
Electric Building,
Butte, Montana 59701.

Attention: Mr. Earl F. Carlson,
Senior Gas Engineer.

Subject: Mercury Instruments Model #840 Full Scallop
Pressure/Volume Recorder with 300# W.P.
American CVMP Positive Displacement Meter.

Special Approval has been granted by the Standards Branch to the Montana Power Company to install a Mercury Model #840 Pressure/Volume Recorder mounted on an American Base Volume Index which in turn will be mounted on an American Type CVMP positive displacement meter. This equipment will be installed at the Montana-Alberta Border Station at Sweetgrass, Montana for the purpose of exporting natural gas to Canadian-Western Natural Gas Company Limited for their system which serves Milk River, Alberta.

Other details of the instruments are:

Mercury PV Recorder - Model 840
Range 0-500 psig

Base Volume Index - BP-0-300

Positive Displacement
Meter - 3.5M
- 300# W.P.

W. J. S. Fraser

W.J.S. Fraser,
Chief, Electricity & Gas
Division.

c.c. Mr. A. Park
Mr. J. Taylor