



OTTAWA, December 2, 1974

Your file    Votre référence

Our file    Notre référence

GL 1147-57/K3-822

SPECIAL APPROVAL

Granted to:        Alberta Natural Gas Company,  
                      Gold Creek Road,  
                      P.O. Box 39,  
                      Cranbrook, B.C.

Attention:         Mr. J. P. Walls,  
                      District Supervisor

Subject:            Kimray Gas Gravitometer

Special Approval has been granted by the Legal Metrology and Laboratory Services Branch to Alberta Natural Gas Company for the use of the above-named instrument in connection with billing for gas sales.

The following specifications apply to the approved unit.

Serial No.	317557
Gas Company No.	60-2
Specific Gravity Range	0.25 to 1.20
Power Supply	110-120 volts A.C.

The gravitometer will be used to spot check the specific gravity of gas recorded on approved field gravitometers. It may be used in any area under the jurisdiction of the Penticton District Office and shall be subject to quarterly re-verification by the Penticton District.

D. L. Smith,  
Chief,  
Electricity and Gas Division

c.c. Mr. R. W. Ward  
      Mr. A. L. Heaven