



SPG-230

OTTAWA, October 21, 1975

Your file Votre référence

Our file Notre référence

GL-1147-57/C6-112

SPECIAL APPROVAL

Granted to: Saskatchewan Power Corporation,
 2025 Victoria Avenue,
 Regina, Saskatchewan.
 S4P OSI

Attention: Mr. J. W. Campbell,
 Manager, Gas Distribution.

Subject: Canadian Meter Company, Model GT-6
 Gas Turbine Meter, 720 psig W.P.

Special Approval has been granted by the Legal Metrology and Laboratory Services Branch to Saskatchewan Power Corporation to install one Canadian Meter Company, Model GT-6 gas turbine meter rated at 720 psig for measurement of natural gas.


Details of the meter and auxiliary attachment are as follows:

1. Turbine Meter

Company number	SPC 29799
Meter body serial number	75-12925
Cartridge serial number	M-9088
Max. rated pressure	720 psig
Normal operating line pressure	200 psig
Normal operating time temperature	60°F
Estimated load	218,000 SCFH

2. Auxiliary Attachment
Canadian Model 100 BVI-CI
Instrument Serial number PP-1-1118
Pressure range 0-300 psig
Temperature range -20 to +140°F
Input shaft capacity 100 cu. ft. per rev.
Atmospheric pressure 13.5 psia
Base pressure 14.65 psia
Base temperature 60°F

The installation is located at Lincoln-McKay,
Saskatoon, Saskatchewan.



D.L. Smith,
Chief,
Electricity & Gas Division.

c.c. Mr. R.A. Kanerva, CMC, Milton
Mr. J. Duff, D.I., Regina
Mr. J. M. Taylor, R.S., Winnipeg
Mr. R.N. Beattie, D.I., Hamilton.
Mr. J. Taylor, Winnipeg
Mr. F. W. Mason, Toronto