

SPG-267

1981 07 15

For the / Pour le

Number / Numéro 66635-R2-25

APPROBATION SPECIALE


SPECIAL APPROVAL

Accordée à:

Granted to:

Union Gas Limited
50 Kiel Drive North
Chatham, Ontario
N7M 5M1

Attention: Mr. F. Boote,
Development Engineer.



(pvt) D. L. Smith, P. Eng.

Chef
Division de l'électricité et du gaz
Direction de la métrologie légale
Edifice des normes
Avenue Holland
Ottawa (Ontario)
K1A 0G9

Chief
Electricity and Gas Division
Legal Metrology Branch
Standards Building
Holland Avenue
Ottawa (Ontario)
K1A 0G9

SPECIAL APPROVAL 267

Subject: Modification to Rockwell T-60 Turbo Gas Meter,
Serial Number 81-004),
Located at the Canadian Gypsum Company,
Hagersville, Ontario.

Special approval has been granted by the Legal Metrology Branch to Union Gas Limited, to outfit the meter described above with a heating cable.

The heating cable is to be installed in the interior of the gear train compartment where it surrounds the change gear assembly. Also, the same cable is wrapped around the exterior of the gear train assembly housing.

This Special Approval does not allow for any modifications to the meter which could affect its accuracy, nor does this approval imply that this modification has been approved for meeting safety requirements.

Furthermore, this Special Approval is limited to the particular meter described above at the designated location.

cc: Mr. W. J. Taggart
District Manager
Electricity and Gas - HAMILTON