



Date: February 21, 2012

File number: 26686-AP-AM-11-0241

MODIFICATION ACCEPTANCE LETTER (MAL) FOR Approved Weighing and Measuring Devices

Type of Device

Static Electronic Multiple Dimension Measuring Device

Purpose

The purpose of this letter is to convey details of modifications to weighing and measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect devices identified in the following Notice(s) of Approval:

Approval Number

AM-5567C

Applicant

Mettler Toledo
1150 Dearborne Drive
Worthington, Ohio, 43085
USA/É.U.

Description of Modifications

- The purpose of this modification is to correct the device description on page 3 of the Notice of Approval. The device uses a visible red laser, not an infra-red laser.

Original copy signed by:
Isabelle Tremblay
Laboratory Manager
Mass Measurement Discipline



Date: 21 février 2012
Dossier: 26686-AP-AM-11-0241

LETTRE D'ACCEPTATION DE MODIFICATION (LAM) pour Appareils de pesage et de mesure approuvés

Type d'Appareil

Appareil électronique de mesure multidimensionnelle statique

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux appareils de pesage et de mesure dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux appareils visés par l'(les) avis d'approbation suivant(s):

Numéro d'approbation

AM-5567C

Requérant

Mettler Toledo
1150 Dearborne Drive
Worthington, Ohio, 43085
USA/É.U.

Descriptions des modifications

- Cette modification vise à faire la correction de la description sur la page 3 de l'avis d'approbation. L'appareil utilise un laser rouge visible, et non un laser infrarouge.

Copie authentique signée par :
Isabelle Tremblay
Gestionnaire du laboratoire
Discipline des masses