

TRADE AND COMMERCE
CANADA

SD-WA.27

STANDARDS DIVISION

OTTAWA, January 7, 1949

NOTIFICATION OF TYPE LISTING

Under the provisions of the appropriate Act the apparatus specified and illustrated herein have been listed as Class A devices, and may be used in Canada in accordance with the conditions applicable to this classification.

Authority: Weights and Measures Act, Chapter 212, R.S. 1927, and Regulations thereunder (P.C. 3459).

Apparatus Listed: Model 4010 and 4016 "Over and Under" Standard type Balance Indicators, manufactured by the Exact Weight Scale Company, Columbus, Ohio.

Rating of Apparatus: N.A.

Application: For attachment to standard beam platform scales.

Conditions: N.A.

Description: The apparatus consists of a long lever which is operated by a draft hand clamped to the scale beam. An adjustable zero marking is provided to compensate for uneven flooring. The zero adjustment is accomplished by a screw and thumb nut combination. The model 4016 indicator is also provided with a quick balancing device. The mechanism of both models is enclosed in a sheet metal housing.

Testing: N.A.

J. L. Stiver,
Chairman, Examining Board.

A. F. Gill,
Director of Standards.

EXACT WEIGHT OVER AND UNDER STANDARD TYPE BALANCE INDICATOR

MODEL #4010

