

TRADE AND COMMERCE
CANADA

SD-WA.37

STANDARDS DIVISION

OTTAWA, January 15, 1949

NOTIFICATION OF TYPE LISTING

Under the provisions of the appropriate Act, the apparatus specified and illustrated herein have been listed as Class C devices, and may be used in Canada in accordance with the conditions applicable to this classification.

Authority: Weights and Measures Act, Chapter 212, R.S. 1927 and Regulations thereunder (P.C. 3459).

Apparatus Listed: Models 38, 62, 62A, 42, and 42A Automatic Packaging Scales, manufactured by J.D. Merrifield & Son, Rocky Ford, Colorado.

Rating of Apparatus: 25 lb., 25 lb., 25 lb., 100 lb., and 100 lb. maximum capacity respectively.

Conditions: As prescribed in Part II, Sections 3 and 4 of P.C. 3459.

Description: The mechanism of these scales incorporates a poise beam supported by flexible links in place of knife-edges. Automatic operation of the weighing bucket is provided by electric circuits and solenoids. Types 38, 62 and 62A are adjustable to various weights of bucket dumps.

Testing: Not applicable, as Class C devices do not require government inspection.

J. L. Stiver,
Chairman, Examining Board.

A. F. Gill,
Director of Standards.

MERRIFIELD PACKAGING SCALE

MODEL 42

