



CANADA

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS DIVISION

SD-WA.45

OTTAWA.....May 26th, 1949.....

NOTIFICATION OF TYPE LISTING

Under the provisions of the appropriate Act the apparatus specified and illustrated herein has been listed as a Class A device, and may be used in Canada in accordance with the conditions stated herein.

Authority: Weights and Measures Act, Chapter 212, R.S. 1927, and Regulations thereunder (P.C. 3459).

Apparatus Listed: Model 1110-A Pre-packaging Scale, manufactured by Toledo Scale Company of Canada Limited, Windsor, Ont.


Rating of Apparatus: 30 lbs. maximum capacity.

Application: Weighing of commodities for pre-packaging purposes.

Conditions: As prescribed in P.C. 3459 with the added mandatory requirement that the scale shall not be used on the retail premises of any business enterprise.

Description: This scale is identical with the Model 1110 Counter-Computing-Platform Scale listed under circular SD-WA.23 (Revision - 1), dated February 25th, 1949, except that it is provided with an indicating scroll for price selection and a detachable manual knob for tare adjustment.


Testing: The standard tests for a Counter-Computing-Platform Cylindrical Chart Scale will apply.


J. L. Stiver,
Acting Director,
Standards Division.

SD-WA.45

TOLEDO PRE-PACKAGING SCALE

MODEL 1110-A



VIEW OF OPERATOR'S SIDE