



SD-WA.63

CANADA

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS DIVISION

OTTAWA, July 7th, 1950.

NOTIFICATION OF TYPE LISTING

Under the provisions of the appropriate Act the apparatus specified and illustrated herein have been listed as Class A devices, and may be used in Canada in accordance with the conditions stated herein.

Authority: Weights and Measures Act, Chapter 212, R.S. 1927, and Regulations thereunder (P.C. 3459).

Apparatus Listed: Models 213-W and 271-W Over and Under Scales, manufactured by the Exact Weight Scale Company, Columbus, Ohio.


Rating of Apparatus: 5 lb. and 12 lb. capacity respectively.

Application: Weighing of commodities for prepackaging purposes.

Conditions: As prescribed in P.C. 3459 with the added mandatory requirement that these scales shall not be used on the retail premises of any business enterprise.

Description: These scales are identical with the models approved under Notification of Type Listing SD-WA.28 of January 7th, 1949, with the exception that an ungraduated tare beam and sliding poise are provided on the opposite side of the scale to the main beam.

Testing: The standard tests for an equal-arm beam type scale will apply and notice should be taken of the requirements of Part VI, Section 28, of P.C. 3459 respecting prepackaging tolerances.


R. W. MacLean,
Director,
Standards Division.

EXACT WEIGHT OVER AND UNDER SCALE

MODEL 213-W

