

TRADE AND COMMERCE  
CANADA

SD-WA.122

## STANDARDS DIVISION

OTTAWA, September 19th, 1952.

## NOTIFICATION OF TYPE LISTING

Under the provisions of the Weights and Measures Act, Chapter 36, S.C. 1951, and Regulations thereunder (P.C.6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

**Apparatus:** Bench type automatic indicating scale equipped with registering beam and fractional poise weights or a counting beam, Models 1108-W10 and 1108-C, manufactured by the Stimpson Scale Company, Louisville, Kentucky, U.S.A. for the Howe Scale Company, Rutland, Vermont, U.S.A.


**Rating:** Both models - 100 lbs. capacity.


**Application:** Weighing of commodities in general trade and for industrial use.

**Conditions:** As prescribed in P.C.6894.

**Description:** A compound lever scale equipped with a 10 lb. x 1 ounce fan shaped chart, 10 lb. capacity beam and sufficient counterpoise weights to weigh to 100 lbs. The load receiving element is a steel plate and has 4 point suspension. The scale is fitted with V type agate bearings and pear shaped pivots. A hydraulic dash-pot supplies damping effect to the registration mechanism. Movement is transferred from the transfer lever to the indicator by means of a steel ribbon and cam. Load resistance is provided by pendulum weight and counterpoise weights.

**Testing:** The standard tests for an automatic indicating scale shall apply.

  
C. S. Phillips,  
Assistant Director (W&M),  
Standards Division

  
R. W. MacLean,  
Director,  
Standards Division.

Reference A-211.

HOWE FAN TYPE COUNTER SCALE

MODEL 1108-W10

