



OFFICE OF THE DIRECTOR

TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

PLEASE QUOTE FILE.....

OTTAWA, September 23rd, 1952.

NOTIFICATION OF TYPE LISTING

Under the provisions of the Weights and Measures Act, Chapter 36, S.C. 1951, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.


Apparatus: Pennsylvania Over and Under Counter Scales for weighing and counting. Models and weighing capacities: C-50 (10 lbs.), C-140 (20 lbs.), C-142 (20 lbs.), C-143 (20 lbs.), C-200 (20 lbs.), manufactured by the National Store Specialty Company, Lancaster, Pa., U.S.A.


Application: Weighing and counting small parts and particularly adapted for industrial use.

Conditions: As prescribed in P.C.6894.

Description: These devices are equipped with a weighing mechanism and side beams of the same design as Model 40 listed in SD-WA.123. Dial registration is limited to a zero graduation which is viewed through a window in the dial housing. The words OVER and UNDER are printed on the chart on the appropriate side of the zero indication. Scales may be equipped with single, double or triple ratio beams for counting and may be supplied with a scoop. Side beams are 10 lbs. capacity by one ounce graduations. Beams are not notched but poises are provided with means for locking in position. Scales are equipped with friction lock type balance ball, a level and adjustable levelling feet.

Testing: Standard tests for an Over and Under Scale with sidebeam shall apply. Ratio of counting device to be checked.



C. S. Phillips,
Assistant Director (W&M),
Standards Division.


R. W. MacLean,
Director,
Standards Division.

Reference A-221.

NATIONAL STORE SPECIALTY WEIGHING & COUNTING SCALE

MODEL 143.



MODEL C-200

