



CANADA

DEPARTMENT OF TRADE AND COMMERCE  
STANDARDS DIVISION

OTTAWA, July 20th, 1953.

## NOTIFICATION OF TYPE LISTING

Under the provisions of the Weights and Measures Act, Chapter 36, S.C. 1951, and Regulations thereunder (P.C.6894), the apparatus specified and illustrated herein has been listed as an approved device and may be used in Canada in accordance with the conditions applicable.

**Apparatus Listed:** Counter cylinder chart prepackaging scale, Model S1P, capacity - 15 lbs. x 1/4 ounces, manufactured by the Globe Slicing Machine Company, Louisville, Kentucky, U.S.A.

**Application:** Weighing of commodities for prepackaging purposes.

**Conditions:** As prescribed in P.C.6894 with added mandatory requirements as follows:

- (1) That the scales be not used on the retail premises of a business enterprise.
- (2) That the scale bears the legend "LEGAL FOR PREPACKAGING USE ONLY AND NOT FOR DIRECT SALES TO CUSTOMER" in a permanent manner located in a position plainly visible to an observer. The letters must be at least 3/16" in height and of a contrasting color to the background.

**Description:** This device is similar in general design and construction to Model S1 (SD-WA.94), except that the chart and zero adjustment are different and a taring device is added.


The chart is cylindrical and in two sections with .1¢ graduations on the left side to .95¢ per lb., and .2¢ graduations on the right side to \$1.80 per lb.

The weight graduations are on the extreme right of the chart viewed from the trader's side and may also be viewed through a window on the customer's side. A beam of light in combination with a 2 $\frac{1}{2}$  power lense projects the weight and money graduations on a ground glass screen which extends across the cylinder casing, the money value lense being moved to the appropriate price per lb. by a small crank projecting from the side of the scale, the taring device is operated by a knob set in the top of the cylinder casing and the zero balance adjustment may be operated by a screw driver through a hole in the base of the scale on the trader's side.

The scale is finished in white porcelain.

**Testing:** The standard tests for a cylinder type counter computing scale shall apply.

  
C. S. Phillips,  
Assistant Director (W&M),  
Standards Division.

  
R. W. MacLean,  
Director,  
Standards Division.

Reference A: 262

GLOBE - STIMPSON PREPACKAGING SCALE

MODEL 81-P

