



TRADE AND COMMERCE  
CANADA

## STANDARDS DIVISION

OTTAWA, September 29th, 1953.

## NOTIFICATION OF TYPE LISTING

Under the provisions of the Weights and Measures Act, Chapter 36, S.C. 1951, and Regulations thereunder (P.C.6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

**Apparatus Listed:** Detecto Platform dial scales, bench and portable floor types; the bench type includes model numbers 2200, 2201, 2202, 2203, 2204, 2205, 2206, 2121, 2124, and 2128; the floor type includes model numbers 2321, 2324, 2326, and 2328.

The above model numbers by the addition of a suffix made up of numbers and letters form catalogue numbers which denote (1) the capacity; (2) whether the scale is or is not equipped with tare and capacity beams; (3) size of platform; (4) if supplied with stand.

These scales are manufactured by Detecto Scales, Inc., Brooklyn, New York, U.S.A.


**Rating of Apparatus:** The capacities of the bench models range from 50 lbs. to a maximum of 2250 lbs. and the capacities of the floor models range from 125 lbs. to 2250 lbs.


**Application:** Weighing of material in wholesale quantities, warehouse weighing and industrial use.

**Conditions:** As prescribed in P.C.6894.

**Description:** These mechanisms are all portable platform dial scales of the same fundamental design. The dial head is identical with that listed in Circular SD-WA.83. The leverage system consists of two T type levers coupled with a transfer lever. Pivots are triangular, pressed into slots and are hardened. Bearings are hardened V type or loop type in the case of the load bearings. The platform is pressed steel sheet with load extensions from the frame which are suspended by 4 loop bearings. The scale is equipped with a dash pot and the main beam is notched if the scale is so equipped.

**Testing:** The standard tests for a platform dial scale shall apply.

  
C. S. Phillips,  
Assistant Director (W&M),  
Standards Division.

  
R. W. MacLean,  
Director,  
Standards Division.

Reference: A-244.


RSI AM-02

SD-WA-189



DETECTO BENCH SCALE  
MEMORANDUM DINA SCART  
ACUMAC  
CATALOGUE NUMBER  
2293 AT BRADYACTS  
NOIZIVIS AT BRADYACTS

PROY. 01983 22930402 AWAITTO



PROY. 01983 22930402

PROY. 01983 22930402