

TRADE AND COMMERCE
CANADA

SD-WA. 211

STANDARDS DIVISION

OTTAWA, February 10th, 1955.

NOTIFICATION OF TYPE LISTING

Computing Counter Scale - Manufactured by
Toledo Scale Company of Canada, Limited,
2462 Howard Avenue, Windsor, Ont.

Under the provisions of the Weights and Measures Act,
Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the
apparatus specified and illustrated herein has been listed as an approved
device and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Computing counter scale, Model 1120 with a capacity
of 30 lbs. x 1 oz.

Application: Weighing commodities in trade.

Conditions: As prescribed in P.C. 6894.


Description: This device is a computing counter scale equipped with
a cylindrical chart and is identical in basic design
and construction to Model 1110 listed on SD-WA. 23,
except for the following minor changes incorporated in
the new model.

1. Florescent lighting has been installed.
2. No customer side lighting.
3. The trader's sighting glass tilts to a
greater angle.
4. A stainless steel platter with a masonite
cover has been installed.
5. The dash pot has been redesigned to give
a quicker stop.
6. The new model is equipped with 4 levelling
feet instead of 3.

The scale casing is finished in white porcelain.

Testing: The standard tests for a computing counter scale shall apply.


G. S. Phillips,
Assistant Director (W&M),
Standards Division.


R. W. MacLean,
Director,
Standards Division.

Reference: A-321.

TOLEDO COMPUTING COUNTER SCALE

MODEL 1120

