



CANADA

SD - WA. 264

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS DIVISION

OTTAWA, March 27th, 1956.

APPROVAL LISTING

Equal arm bench Scale manufactured by
W & T Avery Ltd., Birmingham, England.

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Equal arm scale bench type with semi automatic registering head, Model 1302 BDL with center tower 860 lbs. capacity, Model 1303 BDL with end tower and 112 lbs. capacity.

Application: Weighing commodities in wholesale trade, warehouse and industrial weighing.

Description: These devices are equal arm scales with semi automatic indicating dials which may be of the over and under type with graduations and weight values plus and minus from zero or the charts may be graduated plus from zero only. When minus graduations are used they are printed in a contrasting color to the plus graduations. The registering head mechanism is illustrated on the back of this circular. The leverage system is similar in design to Model 1301 (Circular SD-WA263) except that in Model 1303BDL the registering head actuating mechanism is modified to suit the end tower assembly. Weights are supplied to the capacity of the scale. The scale may be equipped with pan, platter or scoop type load receiving element.

Testing: The standard tests for an equal arm scale shall apply and in addition the chart graduations shall be subjected to tests applying to automatic indicating scales.

C. S. Phillips,
Assistant Director (W&M),
Standards Division.


R. W. MacLean,
Director,
Standards Division.

Reference: A-382 .


SD-WA.264

APPROX BRIDGE SCALE

INDICATING HEAD MECHANISM



MODEL 1500000



MODEL 1500000

