



CANADA

SD-WA. 266

DEPARTMENT OF TRADE AND COMMERCE  
STANDARDS DIVISION

OTTAWA, April 11th, 1956.

APPROVAL LISTING

Automatic Indicating Counter Scale  
Manufactured by W & T Avery Co. Ltd., Birmingham, Eng.

Under the provisions of the Weights and Measures Act, Chapter 292, R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and illustrated herein have been listed as approved devices and may be used in Canada in accordance with the conditions applicable.

Apparatus Listed: Automatic indicating non-computing counter scale, Model 1105CCE with capacities of 1 lb. and 2 lbs.

Application: Weighing commodities in general trade.

Conditions: As prescribed in P.C. 6894.

Description: This device is an automatic indicating counter scale with a fan type chart. The single equal arm lever is loaded at one end to compensate for the weight of the load pan and its support. Movement is transferred to the indicator by means of a cam and lever mechanism and resistance to load is provided by a single pendulum weight. The scale is equipped with a thermostatically controlled dash pot and is enclosed in a pressed steel casing with adjustable legs. A shot pot is located under the load pan.

The load receiving element may be a pan, platter or scoop.

Testing: The standard tests for an automatic indicating counter scale shall apply.

G. S. Phillips,  
Assistant Director, (W&M),  
Standards Division.

R. W. MacLean,  
Director,  
Standards Division.

Reference: A-380

AVEY AUTOMATIC INDICATING SCALE

MODEL 110000E

Registering Hood Mechanism

SD-1A. 266

